

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV112019\
 Data File : VV013686.D
 Acq On : 20 Nov 2019 14:01
 Operator : SY/MD
 Sample : VV1120WBL01
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK58

Quant Time: Nov 21 01:08:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111519WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Nov 21 00:51:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	415089	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	409695	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	175456	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	147644	5.76	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	115.20%
7) Chloroethane-d5	1.58	69	128470	5.61	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	112.20%
11) 1,1-Dichloroethene-d2	2.12	63	199622	4.34	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.80%
20) 2-Butanone-d5	3.95	46	297310	50.14	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.28%
24) Chloroform-d	4.39	84	291445	5.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.80%
26) 1,2-Dichloroethane-d4	5.07	65	148725	5.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.20%
32) Benzene-d6	5.09	84	590740	5.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.40%
36) 1,2-Dichloropropane-d6	6.11	67	174970	5.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.80%
41) Toluene-d8	7.35	98	501340	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.40%
43) trans-1,3-Dichloropropene-	7.66	79	60487	4.66	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.20%
46) 2-Hexanone-d5	8.14	63	260177	52.94	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.88%
57) 1,1,2,2-Tetrachloroethane-	10.25	84	125472	5.14	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	191911	5.68	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV112019\
Data File : VV013686.D
Acq On : 20 Nov 2019 14:01
Operator : SY/MD
Sample : VV1120WBL01
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK58

Quant Time: Nov 21 01:08:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111519WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Nov 21 00:51:34 2019
Response via : Initial Calibration

