

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW112023\
 Data File : VW033029.D
 Acq On : 20 Nov 2023 18:12
 Operator : SY/MD
 Sample : VIBLK252
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK252

Quant Time: Nov 20 23:42:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM112023WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Nov 20 23:40:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	283624	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	281290	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	149628	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	61401	41.405	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.820%
7) Chloroethane-d5	1.532	69	48412	42.448	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.900%
11) 1,1-Dichloroethene-d2	2.060	65	23635	38.952	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.900%
21) 2-Butanone-d5	3.783	46	88393	93.686	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.690%
24) Chloroform-d	4.249	84	156216	42.745	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.500%
26) 1,2-Dichloroethane-d4	4.941	65	102244	44.850	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.700%
32) Benzene-d6	4.960	84	281277	42.940	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.880%
36) 1,2-Dichloropropane-d6	5.989	67	84762	45.632	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.260%
41) Toluene-d8	7.246	98	241535	38.626	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	77.260%#
43) trans-1,3-Dichloroprop...	7.555	79	44942	41.443	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.880%
47) 2-Hexanone-d5	8.024	63	71693	93.270	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.270%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	120652	42.671	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.340%
66) 1,2-Dichlorobenzene-d4	11.561	152	126013	47.227	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.460%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112023\
Data File : VV033029.D
Acq On : 20 Nov 2023 18:12
Operator : SY/MD
Sample : VIBLK252
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK252

Quant Time: Nov 20 23:42:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM112023WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Nov 20 23:40:05 2023
Response via : Initial Calibration

