

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV112118\
 Data File : VV008690.D
 Acq On : 21 Nov 2018 10:34
 Operator : SY/MD
 Sample : VV1121WBL01
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Nov 22 00:44:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Nov 22 00:36:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	171193	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	153258	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	68377	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	37845	4.98	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.60%
7) Chloroethane-d5	1.59	69	29942	4.78	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.60%
11) 1,1-Dichloroethene-d2	2.13	63	55619	3.60	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.00%
20) 2-Butanone-d5	3.98	46	110908	40.76	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	81.52%
24) Chloroform-d	4.40	84	90450	4.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.00%
26) 1,2-Dichloroethane-d4	5.09	65	47205	4.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.20%
32) Benzene-d6	5.10	84	191846	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
36) 1,2-Dichloropropane-d6	6.12	67	56590	4.39	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.80%
41) Toluene-d8	7.36	98	170819	4.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.00%
43) trans-1,3-Dichloropropene-	7.67	79	18393	3.74	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.80%
46) 2-Hexanone-d5	8.14	63	78360	35.18	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	70.36%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	37608	4.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	55515	4.50	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.00%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV112118\
Data File : VV008690.D
Acq On : 21 Nov 2018 10:34
Operator : SY/MD
Sample : VV1121WBL01
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Nov 22 00:44:08 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Nov 22 00:36:45 2018
Response via : Initial Calibration

