

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV112119\
 Data File : VV013709.D
 Acq On : 21 Nov 2019 13:31
 Operator : SY/MD
 Sample : K5911-11
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GAQL4

Quant Time: Nov 22 04:40:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111519WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 22 03:55:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	345963	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	347942	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	153175	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	127681	5.97	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	119.40%
7) Chloroethane-d5	1.58	69	118585	6.22	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	124.40%
11) 1,1-Dichloroethene-d2	2.12	63	176383	4.60	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	92.00%
20) 2-Butanone-d5	3.96	46	294228	59.54	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	119.08%
24) Chloroform-d	4.40	84	271025	5.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	114.80%
26) 1,2-Dichloroethane-d4	5.08	65	136964	6.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	120.60%
32) Benzene-d6	5.09	84	543225	5.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.00%
36) 1,2-Dichloropropane-d6	6.12	67	164866	5.75	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	115.00%
41) Toluene-d8	7.36	98	476831	5.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.80%
43) trans-1,3-Dichloropropene-	7.66	79	56782	5.15	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	103.00%
46) 2-Hexanone-d5	8.14	63	209257	50.13	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.26%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	109045	5.26	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.20%
64) 1,2-Dichlorobenzene-d4	11.67	152	190114	6.45	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	129.00%#

Target Compounds					Ovalue
42) Toluene	7.43	91	26559	0.236 ug/L	98
52) Ethylbenzene	9.05	91	8588	0.072 ug/L	97
53) m,p-xylene	9.18	106	9390	0.208 ug/L	90
54) o-xylene	9.58	106	3854	0.089 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV112119\
Data File : VV013709.D
Acq On : 21 Nov 2019 13:31
Operator : SY/MD
Sample : K5911-11
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GAQL4

Quant Time: Nov 22 04:40:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR111519WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 22 03:55:18 2019
Response via : Initial Calibration

