

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112124\
 Data File : VV038233.D
 Acq On : 21 Nov 2024 11:57
 Operator : SY/MD
 Sample : P4883-14
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG3L3

Quant Time: Nov 22 01:43:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 22 01:41:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	237738	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	257844	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	107181	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	68133	3.739	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.800%
7) Chloroethane-d5	1.529	69	68866	4.672	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.400%
11) 1,1-Dichloroethene-d2	2.053	65	30890	3.939	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.800%
20) 2-Butanone-d5	3.780	46	188283	51.651	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.300%
24) Chloroform-d	4.243	84	161967	4.580	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
26) 1,2-Dichloroethane-d4	4.937	65	82325	5.188	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.800%
32) Benzene-d6	4.954	84	279661	4.196	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.000%
36) 1,2-Dichloropropane-d6	5.982	67	94916	4.629	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.600%
41) Toluene-d8	7.240	98	221708	3.595	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.000%
43) trans-1,3-Dichloroprop...	7.558	79	28505	3.739	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.800%
46) 2-Hexanone-d5	8.018	63	155954	52.623	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.240%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	83768	4.812	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.200%
66) 1,2-Dichlorobenzene-d4	11.555	152	93198	5.103	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.000%
Target Compounds						
2) Dichlorodifluoromethane	1.105	85	9890	0.440	ug/L	100
10) 1,1,2-Trichloro-1,2,2-...	2.060	101	5701	0.320	ug/L	92
13) Acetone	2.111	43	11611	4.307	ug/L	91
16) Methylene chloride	2.442	84	18568	0.959	ug/L	92
22) cis-1,2-Dichloroethene	3.809	96	17381	0.940	ug/L	95
29) 1,1,1-Trichloroethane	4.500	97	27753	0.897	ug/L	97
34) Trichloroethene	5.825	95	201342	10.198	ug/L	97
42) Toluene	7.326	91	11972	0.156	ug/L	94
47) Tetrachloroethene	7.899	164	115700	7.293	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112124\
Data File : VV038233.D
Acq On : 21 Nov 2024 11:57
Operator : SY/MD
Sample : P4883-14
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG3L3

Quant Time: Nov 22 01:43:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 22 01:41:27 2024
Response via : Initial Calibration

