

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV112219\
 Data File : VV013737.D
 Acq On : 22 Nov 2019 16:36
 Operator : SY/MD
 Sample : K5941-07
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 B0AB9

Quant Time: Nov 23 02:45:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Nov 23 01:35:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	338820	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	329427	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	150890	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	113365	5.41	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	108.20%
7) Chloroethane-d5	1.58	69	103519	5.81	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	116.20%
11) 1,1-Dichloroethene-d2	2.13	63	171996	4.87	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	97.40%
20) 2-Butanone-d5	3.92	46	245193	56.24	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.48%
24) Chloroform-d	4.40	84	226913	5.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.40%
26) 1,2-Dichloroethane-d4	5.08	65	114701	5.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	114.00%
32) Benzene-d6	5.10	84	445276	5.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.80%
36) 1,2-Dichloropropane-d6	6.11	67	137932	5.45	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.00%
41) Toluene-d8	7.36	98	389383	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.60%
43) trans-1,3-Dichloropropene-	7.66	79	47271	4.80	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.00%
46) 2-Hexanone-d5	8.13	63	175434	48.24	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.48%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	95607	5.44	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	156419	5.87	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.40%

Target Compounds

					Ovalue
3) Chloromethane	1.25	50	7341	0.370 ug/L	99
12) 1,1-Dichloroethene	2.14	96	42411	2.603 ug/L #	75
14) Carbon disulfide	2.32	76	14731	0.392 ug/L	98
16) Methylene chloride	2.53	84	4693	0.210 ug/L	96
17) Methyl tert-butyl Ether	2.80	73	9539	0.214 ug/L #	91
19) 1,1-Dichloroethane	3.23	63	107958	2.940 ug/L	98
22) cis-1,2-Dichloroethene	3.96	96	44247	2.051 ug/L	98
34) Trichloroethene	5.96	95	259619	11.647 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV112219\
Data File : VV013737.D
Acq On : 22 Nov 2019 16:36
Operator : SY/MD
Sample : K5941-07
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
B0AB9

Quant Time: Nov 23 02:45:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Nov 23 01:35:18 2019
Response via : Initial Calibration

