

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112221\
 Data File : VV023654.D
 Acq On : 22 Nov 2021 11:41
 Operator : SY/MD
 Sample : VSTD0.131
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.1231

Quant Time: Nov 23 01:52:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM112221.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Nov 23 01:51:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	11896	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.864	117	11640	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.244	152	6053	0.500	ug/L	-0.02
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	1092	0.066	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.043	65	696	0.072	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.069	67	859	0.069	ug/L	0.00
8) Toluene-d8	7.332	98	2122	0.066	ug/L	0.02
10) 1,1,2,2-Tetrachloroeth...	10.216	84	502	0.063	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	1533	0.103	ug/L	92
6) Trichloroethene	5.925	95	1082	0.094	ug/L	95
9) 1,2-Dibromoethane	8.369	107	661	0.098	ug/L #	90
12) 1,2,3-Trichloropropane	10.267	75	574	0.098	ug/L	97
13) 1,2-Dibromo-3-chloropr...	12.427	75	94	0.080	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112221\
Data File : VV023654.D
Acq On : 22 Nov 2021 11:41
Operator : SY/MD
Sample : VSTD0.131
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.1231

Quant Time: Nov 23 01:52:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM112221.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Nov 23 01:51:12 2021
Response via : Initial Calibration

