

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW112221\  
 Data File : VV023660.D  
 Acq On : 22 Nov 2021 14:38  
 Operator : SY/MD  
 Sample : M4642-03  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 X3803

Quant Time: Nov 23 02:10:15 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVSIM112221.M  
 Quant Title : TRACE VOA SOM01.0  
 QLast Update : Tue Nov 23 02:04:29 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|--------|----------|
| -----                         |        |       |          |          |        |          |
| Internal Standards            |        |       |          |          |        |          |
| 1) 1,4-Difluorobenzene        | 5.639  | 114   | 11866    | 0.500    | ug/L   | 0.00     |
| 5) Chlorobenzene-d5           | 8.863  | 117   | 11340    | 0.500    | ug/L   | 0.00     |
| 11) 1,4-Dichlorobenzene-d4    | 11.244 | 152   | 5906     | 0.500    | ug/L   | -0.02    |
| System Monitoring Compounds   |        |       |          |          |        |          |
| 2) Vinyl Chloride-d3          | 1.303  | 65    | 6094     | 0.506    | ug/L   | 0.00     |
| Spiked Amount                 | 0.500  | Range | 40 - 130 | Recovery | =      | 102.000% |
| 4) 1,2-Dichloroethane-d4      | 5.042  | 65    | 4594     | 0.561    | ug/L   | 0.00     |
| Spiked Amount                 | 0.500  | Range | 70 - 130 | Recovery | =      | 112.000% |
| 7) 1,2-Dichloropropane-d6     | 6.068  | 67    | 5622     | 0.571    | ug/L   | 0.00     |
| Spiked Amount                 | 0.500  | Range | 60 - 140 | Recovery | =      | 114.000% |
| 8) Toluene-d8                 | 7.317  | 98    | 14527    | 0.570    | ug/L   | 0.00     |
| Spiked Amount                 | 0.500  | Range | 70 - 130 | Recovery | =      | 114.000% |
| 10) 1,1,2,2-Tetrachloroeth... | 10.216 | 84    | 3423     | 0.572    | ug/L   | 0.00     |
| Spiked Amount                 | 0.500  | Range | 65 - 120 | Recovery | =      | 114.000% |
| Target Compounds              |        |       |          |          |        |          |
|                               |        |       |          |          | Qvalue |          |
| 3) Vinyl chloride             | 1.310  | 62    | 2338     | 0.176    | ug/L   | 92       |
| 6) Trichloroethene            | 5.925  | 95    | 9306     | 0.974    | ug/L   | 99       |
| 9) 1,2-Dibromoethane          | 8.353  | 107   | 9455     | 1.640    | ug/L # | 97       |
| 12) 1,2,3-Trichloropropane    | 10.267 | 75    | 2600     | 0.519    | ug/L   | 99       |
| 13) 1,2-Dibromo-3-chloropr... | 12.409 | 75    | 1514     | 1.688    | ug/L # | 72       |
| -----                         |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV112221\  
Data File : VV023660.D  
Acq On : 22 Nov 2021 14:38  
Operator : SY/MD  
Sample : M4642-03  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
X3803

Quant Time: Nov 23 02:10:15 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVSIM112221.M  
Quant Title : TRACE VOA SOM01.0  
QLast Update : Tue Nov 23 02:04:29 2021  
Response via : Initial Calibration

