

Method Path : Z:\voasrv\HPCHEM1\MSVOA_V\Method\
 Method File : SFAMVSIM112221.M
 Title : TRACE VOA SOM01.0
 Last Update : Tue Nov 23 02:04:29 2021
 Response Via : Initial Calibration

Calibration Files

0.5 =VV023655.D 0.05=VV023653.D 0.1 =VV023654.D 1.0 =VV023656.D 2 =VV023657.D

Compound		0.5	0.05	0.1	1.0	2	Avg	%RSD

1) I	1,4-Difluorobenzene	-----ISTD-----						
2) S	Vinyl Chloride-d3	0.517	0.408	0.459	0.558	0.598	0.508	15.00
3) T	Vinyl chloride	0.541	0.422	0.644	0.579	0.613	0.560	15.38
4) S	1,2-Dichloroetha...	0.384	0.253	0.293	0.396	0.400	0.345	19.64
5) I	Chlorobenzene-d5	-----ISTD-----						
6) T	Trichloroethene	0.441	0.329	0.465	0.432	0.440	0.421	12.61
7) S	1,2-Dichloroprop...	0.493	0.335	0.369	0.484	0.491	0.434	17.56
8) S	Toluene-d8	1.258	0.850	0.912	1.278	1.320	1.124	19.91
9) T	1,2-Dibromoethane	0.273	0.188	0.284	0.263	0.263	0.254	14.87
10) S	1,1,2,2-Tetrachl...	0.306	0.201	0.216	0.298	0.298	0.264	19.34
11) I	1,4-Dichlorobenzen...	-----ISTD-----						
12)	1,2,3-Trichlorop...	0.450	0.300	0.474	0.449	0.448	0.424	16.56
13) T	1,2-Dibromo-3-ch...	0.078	0.067	0.078	0.078	0.078	0.076	6.28

(#) = Out of Range