

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112221\
 Data File : VV023657.D
 Acq On : 22 Nov 2021 12:53
 Operator : SY/MD
 Sample : VSTD00234
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD002234

Quant Time: Nov 23 01:52:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM112221.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Nov 23 01:51:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	12146	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	11954	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	6013	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.310	65	29073	1.721	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	19445	1.975	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.068	67	23470	1.845	ug/L	0.00
8) Toluene-d8	7.317	98	63102	1.907	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	14268	1.752	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	29800	1.953	ug/L	96
6) Trichloroethene	5.925	95	21029	1.770	ug/L	99
9) 1,2-Dibromoethane	8.353	107	12575	1.810	ug/L #	98
12) 1,2,3-Trichloropropane	10.267	75	10785	1.860	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.426	75	1868	1.606	ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112221\
Data File : VV023657.D
Acq On : 22 Nov 2021 12:53
Operator : SY/MD
Sample : VSTD00234
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD002234

Quant Time: Nov 23 01:52:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM112221.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Nov 23 01:51:12 2021
Response via : Initial Calibration

