

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV112221\  
 Data File : VV023659.D  
 Acq On : 22 Nov 2021 14:09  
 Operator : SY/MD  
 Sample : VV1122WBL01  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VBLK207

Quant Time: Nov 23 02:10:09 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVSIM112221.M  
 Quant Title : TRACE VOA SOM01.0  
 QLast Update : Tue Nov 23 02:04:29 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|-------|----------|
| -----                         |        |       |          |          |       |          |
| Internal Standards            |        |       |          |          |       |          |
| 1) 1,4-Difluorobenzene        | 5.615  | 114   | 11904    | 0.500    | ug/L  | -0.02    |
| 5) Chlorobenzene-d5           | 8.863  | 117   | 11451    | 0.500    | ug/L  | 0.00     |
| 11) 1,4-Dichlorobenzene-d4    | 11.244 | 152   | 5989     | 0.500    | ug/L  | -0.02    |
| System Monitoring Compounds   |        |       |          |          |       |          |
| 2) Vinyl Chloride-d3          | 1.303  | 65    | 5349     | 0.442    | ug/L  | 0.00     |
| Spiked Amount                 | 0.500  | Range | 40 - 130 | Recovery | =     | 88.000%  |
| 4) 1,2-Dichloroethane-d4      | 5.042  | 65    | 4740     | 0.577    | ug/L  | 0.00     |
| Spiked Amount                 | 0.500  | Range | 70 - 130 | Recovery | =     | 116.000% |
| 7) 1,2-Dichloropropane-d6     | 6.068  | 67    | 5814     | 0.585    | ug/L  | 0.00     |
| Spiked Amount                 | 0.500  | Range | 60 - 140 | Recovery | =     | 116.000% |
| 8) Toluene-d8                 | 7.317  | 98    | 14797    | 0.575    | ug/L  | 0.00     |
| Spiked Amount                 | 0.500  | Range | 70 - 130 | Recovery | =     | 116.000% |
| 10) 1,1,2,2-Tetrachloroeth... | 10.216 | 84    | 3494     | 0.578    | ug/L  | 0.00     |
| Spiked Amount                 | 0.500  | Range | 65 - 120 | Recovery | =     | 116.000% |

Target Compounds Qvalue

-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV112221\  
Data File : VV023659.D  
Acq On : 22 Nov 2021 14:09  
Operator : SY/MD  
Sample : VV1122WBL01  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 8 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VBLK207

Quant Time: Nov 23 02:10:09 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVSIM112221.M  
Quant Title : TRACE VOA SOM01.0  
QLast Update : Tue Nov 23 02:04:29 2021  
Response via : Initial Calibration

