

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV112320\
 Data File : VV019427.D
 Acq On : 23 Nov 2020 17:52
 Operator : SY/MD
 Sample : L4794-17
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BG8

Quant Time: Nov 24 11:34:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM112320WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 24 05:27:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	359324	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	371529	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	175662	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	172935	48.19	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.38%
7) Chloroethane-d5	1.57	69	143584	49.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.32%
11) 1,1-Dichloroethene-d2	2.12	63	243122	38.06	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.12%
21) 2-Butanone-d5	3.89	46	219991	100.58	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.58%
24) Chloroform-d	4.37	84	279315	48.14	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.28%
26) 1,2-Dichloroethane-d4	5.05	65	193414	50.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.86%
32) Benzene-d6	5.07	84	548231	49.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.82%
36) 1,2-Dichloropropane-d6	6.09	67	183153	50.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.22%
41) Toluene-d8	7.33	98	470923	47.86	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.72%
43) trans-1,3-Dichloropropene-	7.64	79	85963	48.69	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.38%
47) 2-Hexanone-d5	8.11	63	119562	83.03	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	83.03%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	194435	41.10	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.20%
64) 1,2-Dichlorobenzene-d4	11.65	152	188550	55.27	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.54%

Target Compounds				Ovalue
48) 2-Hexanone	8.16	43	25289	7.384 ug/L # 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV112320\
Data File : VV019427.D
Acq On : 23 Nov 2020 17:52
Operator : SY/MD
Sample : L4794-17
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BG8

Quant Time: Nov 24 11:34:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM112320WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 24 05:27:09 2020
Response via : Initial Calibration

