

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112321\
 Data File : VV023670.D
 Acq On : 23 Nov 2021 14:56
 Operator : SY/MD
 Sample : VV1123WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK261

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/24/2021
 Supervised By : Mahesh Dadoda 11/26/2021

Quant Time: Nov 24 04:45:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 24 04:42:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	137745	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	138930	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	64333	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	57041	5.044	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 100.800%		
7) Chloroethane-d5	1.568	69	45989	5.174	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 103.400%		
11) 1,1-Dichloroethene-d2	2.111	63	77899	3.909	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 78.200%		
20) 2-Butanone-d5	3.908	46	75760m	55.732	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 111.460%		
24) Chloroform-d	4.349	84	99927	5.075	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 101.400%		
26) 1,2-Dichloroethane-d4	5.034	65	49539	5.385	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 107.800%		
32) Benzene-d6	5.050	84	198196	5.237	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 104.800%		
36) 1,2-Dichloropropane-d6	6.069	67	56388	5.315	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 106.200%		
41) Toluene-d8	7.317	98	176166	4.982	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 99.600%		
43) trans-1,3-Dichloroprop...	7.625	79	21446	5.015	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 100.200%		
46) 2-Hexanone-d5	8.092	63	67507	47.509	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 95.020%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	37376	4.895	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 98.000%		
66) 1,2-Dichlorobenzene-d4	11.625	152	65990	5.802	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 116.000%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112321\
Data File : VV023670.D
Acq On : 23 Nov 2021 14:56
Operator : SY/MD
Sample : VV1123WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK261

Manual Integrations
APPROVED

Reviewed By : John Carlone 11/24/2021
Supervised By : Mahesh Dadoda 11/26/2021

Quant Time: Nov 24 04:45:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 24 04:42:45 2021
Response via : Initial Calibration

