

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW112321\
 Data File : VW023672.D
 Acq On : 23 Nov 2021 15:44
 Operator : SY/MD
 Sample : M4723-04
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0G06

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/24/2021
 Supervised By : Mahesh Dadoda 11/26/2021

Quant Time: Nov 24 04:59:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 24 04:42:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	136050	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	136154	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	63893	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	55418	4.962	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	99.200%	
7) Chloroethane-d5	1.568	69	44399	5.057	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	101.200%	
11) 1,1-Dichloroethene-d2	2.108	63	76711	3.897	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	78.000%	
20) 2-Butanone-d5	3.905	46	76774	57.181	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	114.360%	
24) Chloroform-d	4.349	84	98600	5.070	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	101.400%	
26) 1,2-Dichloroethane-d4	5.034	65	47529	5.231	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	104.600%	
32) Benzene-d6	5.053	84	193018	5.204	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	104.000%	
36) 1,2-Dichloropropane-d6	6.069	67	55886	5.375	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	107.400%	
41) Toluene-d8	7.317	98	169152	4.881	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.600%	
43) trans-1,3-Dichloroprop...	7.625	79	20319	4.848	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	97.000%	
46) 2-Hexanone-d5	8.092	63	65548	47.071	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	94.140%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	35709	4.772	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	95.400%	
66) 1,2-Dichlorobenzene-d4	11.625	152	65632	5.810	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	116.200%	
Target Compounds					Qvalue	
13) Acetone	2.195	43	5302m	4.389	ug/L	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112321\
Data File : VV023672.D
Acq On : 23 Nov 2021 15:44
Operator : SY/MD
Sample : M4723-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0G06

Manual Integrations
APPROVED

Reviewed By : John Carlone 11/24/2021
Supervised By : Mahesh Dadoda 11/26/2021

Quant Time: Nov 24 04:59:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 24 04:42:45 2021
Response via : Initial Calibration

