

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW112421\
 Data File : VW023704.D
 Acq On : 24 Nov 2021 16:40
 Operator : SY/MD
 Sample : M4821-06
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H4655

Quant Time: Nov 26 01:54:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 26 01:51:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	137690	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	133754	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	64023	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	41325	3.656	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.200%
7) Chloroethane-d5	1.568	69	35018	3.941	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.800%
11) 1,1-Dichloroethene-d2	2.108	63	58383	2.931	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	58.600%#
20) 2-Butanone-d5	3.899	46	81298	59.830	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	119.660%
24) Chloroform-d	4.349	84	78297	3.978	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.600%
26) 1,2-Dichloroethane-d4	5.034	65	42359	4.607	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.200%
32) Benzene-d6	5.050	84	158213	4.342	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.800%
36) 1,2-Dichloropropane-d6	6.069	67	46289	4.532	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.600%
41) Toluene-d8	7.317	98	134937	3.964	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.200%
43) trans-1,3-Dichloroprop...	7.625	79	17678	4.294	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.800%
46) 2-Hexanone-d5	8.088	63	71213	52.057	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.120%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	33568	4.567	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.400%
66) 1,2-Dichlorobenzene-d4	11.625	152	53845	4.757	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.200%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.134	85	1706	0.131	ug/L	98
17) Methyl tert-butyl Ether	2.767	73	495110	26.189	ug/L	99
19) 1,1-Dichloroethane	3.195	63	1971	0.112	ug/L	97
22) cis-1,2-Dichloroethene	3.918	96	7475	0.742	ug/L	100
25) Chloroform	4.381	83	7342	0.373	ug/L	97
34) Trichloroethene	5.921	95	8428	0.825	ug/L	100
47) Tetrachloroethene	7.976	164	89995	9.679	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112421\
Data File : VV023704.D
Acq On : 24 Nov 2021 16:40
Operator : SY/MD
Sample : M4821-06
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H4655

Quant Time: Nov 26 01:54:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 26 01:51:50 2021
Response via : Initial Calibration

