

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112421\
 Data File : VV023702.D
 Acq On : 24 Nov 2021 15:52
 Operator : SY/MD
 Sample : M4723-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Nov 26 01:54:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 26 01:51:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	131995	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	132172	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	60875	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	45416	4.191	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.800%
7) Chloroethane-d5	1.568	69	37966	4.457	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.200%
11) 1,1-Dichloroethene-d2	2.108	63	62082	3.251	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.000%
20) 2-Butanone-d5	3.902	46	59007	45.299	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.600%
24) Chloroform-d	4.349	84	87553	4.640	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.800%
26) 1,2-Dichloroethane-d4	5.034	65	43027	4.881	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.600%
32) Benzene-d6	5.050	84	171179	4.755	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.000%
36) 1,2-Dichloropropane-d6	6.069	67	50676	5.020	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.400%
41) Toluene-d8	7.317	98	147190	4.375	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.600%
43) trans-1,3-Dichloroprop...	7.625	79	18606	4.573	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.400%
46) 2-Hexanone-d5	8.092	63	73013	54.012	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.020%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	36196	4.983	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.600%
66) 1,2-Dichlorobenzene-d4	11.625	152	59084	5.490	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112421\
Data File : VV023702.D
Acq On : 24 Nov 2021 15:52
Operator : SY/MD
Sample : M4723-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Nov 26 01:54:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 26 01:51:50 2021
Response via : Initial Calibration

