

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV112519\
 Data File : VV013764.D
 Acq On : 25 Nov 2019 14:35
 Operator : SY/MD
 Sample : K5910-23
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Nov 26 06:28:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Nov 26 04:36:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	351410	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	362163	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	167185	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	102578	4.72	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.40%
7) Chloroethane-d5	1.58	69	98868	5.35	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.00%
11) 1,1-Dichloroethene-d2	2.13	63	145452	3.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.40%
20) 2-Butanone-d5	3.95	46	283428	62.68	ug/L	0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	125.36%
24) Chloroform-d	4.40	84	227635	5.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.80%
26) 1,2-Dichloroethane-d4	5.08	65	116592	5.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.80%
32) Benzene-d6	5.10	84	444836	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.20%
36) 1,2-Dichloropropane-d6	6.11	67	136931	4.92	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.40%
41) Toluene-d8	7.36	98	384331	4.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.60%
43) trans-1,3-Dichloropropene-	7.66	79	47974	4.43	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.60%
46) 2-Hexanone-d5	8.14	63	223126	55.81	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	111.62%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	107896	5.59	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	111.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	165968	5.62	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV112519\
Data File : VV013764.D
Acq On : 25 Nov 2019 14:35
Operator : SY/MD
Sample : K5910-23
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Nov 26 06:28:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Nov 26 04:36:55 2019
Response via : Initial Calibration

