

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV112519\
 Data File : VV013765.D
 Acq On : 25 Nov 2019 14:59
 Operator : SY/MD
 Sample : K5941-08DL 4X
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 B0AC0DL

Quant Time: Nov 26 06:44:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Nov 26 04:36:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	364131	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	356377	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	164091	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	100844	4.47	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.40%
7) Chloroethane-d5	1.58	69	90091	4.71	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.20%
11) 1,1-Dichloroethene-d2	2.13	63	168744	4.45	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	89.00%
20) 2-Butanone-d5	3.95	46	247126	52.75	ug/L	0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.50%
24) Chloroform-d	4.40	84	216349	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.20%
26) 1,2-Dichloroethane-d4	5.08	65	107808	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.80%
32) Benzene-d6	5.09	84	423415	4.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.20%
36) 1,2-Dichloropropane-d6	6.11	67	129635	4.73	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.60%
41) Toluene-d8	7.36	98	354991	4.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.20%
43) trans-1,3-Dichloropropene-	7.66	79	44508	4.17	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.40%
46) 2-Hexanone-d5	8.14	63	183546	46.66	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.32%
57) 1,1,2,2-Tetrachloroethane-	10.25	84	98158	5.16	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.20%
64) 1,2-Dichlorobenzene-d4	11.67	152	153838	5.31	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.20%

Target Compounds

					Ovalue
12) 1,1-Dichloroethene	2.14	96	61945	3.538 ug/L	99
13) Acetone	2.21	43	6514	1.988 ug/L #	64
14) Carbon disulfide	2.32	76	4767	0.118 ug/L	98
19) 1,1-Dichloroethane	3.23	63	300269	7.608 ug/L	99
22) cis-1,2-Dichloroethene	3.96	96	4673	0.202 ug/L	90
34) Trichloroethene	5.96	95	95311	3.952 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV112519\
Data File : VV013765.D
Acq On : 25 Nov 2019 14:59
Operator : SY/MD
Sample : K5941-08DL 4X
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
B0AC0DL

Quant Time: Nov 26 06:44:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Nov 26 04:36:55 2019
Response via : Initial Calibration

