

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV112520\
 Data File : VV019438.D
 Acq On : 25 Nov 2020 10:25
 Operator : SY/MD
 Sample : VV1125MBL01
 Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK88

Quant Time: Nov 26 00:28:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM112320WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Nov 26 00:24:16 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	385377	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	389256	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	184391	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	152104	39.52	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	79.04%
7) Chloroethane-d5	1.58	69	134310	42.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	85.74%
11) 1,1-Dichloroethene-d2	2.12	63	228320	33.33	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.66%
21) 2-Butanone-d5	3.90	46	216976	92.49	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.49%
24) Chloroform-d	4.37	84	267781	43.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.08%
26) 1,2-Dichloroethane-d4	5.05	65	184272	45.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.48%
32) Benzene-d6	5.07	84	510910	44.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.78%
36) 1,2-Dichloropropane-d6	6.09	67	176526	46.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.20%
41) Toluene-d8	7.33	98	436070	42.30	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.60%
43) trans-1,3-Dichloropropene-	7.64	79	80467	43.51	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.02%
47) 2-Hexanone-d5	8.11	63	119752	79.37	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	79.37%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	199246	40.20	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	80.40%
64) 1,2-Dichlorobenzene-d4	11.65	152	172213	48.09	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.18%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV112520\
Data File : VV019438.D
Acq On : 25 Nov 2020 10:25
Operator : SY/MD
Sample : VV1125MBL01
Misc : 5.0µL/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK88

Quant Time: Nov 26 00:28:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM112320WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Nov 26 00:24:16 2020
Response via : Initial Calibration

