

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW112522\  
 Data File : VW029345.D  
 Acq On : 25 Nov 2022 10:38  
 Operator : SY/MD  
 Sample : N5760-09  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VHBLK001

Quant Time: Nov 25 21:29:56 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR112322WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Fri Nov 25 21:27:54 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|-------|----------|
| -----                         |        |       |          |          |       |          |
| Internal Standards            |        |       |          |          |       |          |
| 1) 1,4-Difluorobenzene        | 5.609  | 114   | 283434   | 5.000    | ug/L  | 0.00     |
| 28) Chlorobenzene-d5          | 8.844  | 117   | 248484   | 5.000    | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.239 | 152   | 122524   | 5.000    | ug/L  | 0.00     |
| System Monitoring Compounds   |        |       |          |          |       |          |
| 4) Vinyl Chloride-d3          | 1.304  | 65    | 84423    | 4.840    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =     | 96.800%  |
| 7) Chloroethane-d5            | 1.565  | 69    | 64223    | 5.030    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =     | 100.600% |
| 11) 1,1-Dichloroethene-d2     | 2.105  | 63    | 102446   | 3.670    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =     | 73.400%  |
| 20) 2-Butanone-d5             | 3.899  | 46    | 120605   | 41.332   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =     | 82.660%  |
| 24) Chloroform-d              | 4.339  | 84    | 150825   | 4.844    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 96.800%  |
| 26) 1,2-Dichloroethane-d4     | 5.027  | 65    | 71309    | 5.206    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 104.200% |
| 32) Benzene-d6                | 5.043  | 84    | 339574   | 4.945    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 98.800%  |
| 36) 1,2-Dichloropropane-d6    | 6.063  | 67    | 99060    | 5.133    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =     | 102.600% |
| 41) Toluene-d8                | 7.307  | 98    | 312307   | 4.818    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 96.400%  |
| 43) trans-1,3-Dichloroprop... | 7.616  | 79    | 36287    | 4.877    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =     | 97.600%  |
| 46) 2-Hexanone-d5             | 8.085  | 63    | 123736   | 46.657   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =     | 93.320%  |
| 56) 1,1,2,2-Tetrachloroeth... | 10.207 | 84    | 60469    | 5.078    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =     | 101.600% |
| 66) 1,2-Dichlorobenzene-d4    | 11.616 | 152   | 110285   | 5.152    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =     | 103.000% |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV112522\  
Data File : VV029345.D  
Acq On : 25 Nov 2022 10:38  
Operator : SY/MD  
Sample : N5760-09  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VHBLK001

Quant Time: Nov 25 21:29:56 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR112322WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Fri Nov 25 21:27:54 2022  
Response via : Initial Calibration

