

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW112522\
 Data File : VW029347.D
 Acq On : 25 Nov 2022 11:27
 Operator : SY/MD
 Sample : N5725-09DL 20X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BHB65DL

Quant Time: Nov 25 21:30:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 25 21:27:54 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	272864	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	239634	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	125259	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	91205	5.431	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 108.600%		
7) Chloroethane-d5	1.561	69	67078	5.457	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 109.200%		
11) 1,1-Dichloroethene-d2	2.102	63	110085	4.097	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 82.000%		
20) 2-Butanone-d5	3.889	46	156675	55.774	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 111.540%		
24) Chloroform-d	4.339	84	158704	5.295	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 105.800%		
26) 1,2-Dichloroethane-d4	5.024	65	74300	5.634	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 112.600%		
32) Benzene-d6	5.040	84	348535	5.263	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 105.200%		
36) 1,2-Dichloropropane-d6	6.060	67	101707	5.465	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 109.400%		
41) Toluene-d8	7.307	98	326750	5.227	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 104.600%		
43) trans-1,3-Dichloroprop...	7.616	79	37809	5.270	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 105.400%		
46) 2-Hexanone-d5	8.085	63	136690	53.445	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 106.900%		
56) 1,1,2,2-Tetrachloroeth...	10.207	84	65700	5.721	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 114.400%		
66) 1,2-Dichlorobenzene-d4	11.612	152	121627	5.558	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 111.200%		
Target Compounds						
12) 1,1-Dichloroethene	2.111	96	10523	0.750	ug/L	Qvalue # 1
16) Methylene chloride	2.500	84	8486	0.396	ug/L	98
19) 1,1-Dichloroethane	3.182	63	36339	1.254	ug/L	98
22) cis-1,2-Dichloroethene	3.902	96	133713	7.260	ug/L	100
29) 1,1,1-Trichloroethane	4.597	97	42411	1.539	ug/L	98
34) Trichloroethene	5.908	95	27691	1.515	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112522\
Data File : VV029347.D
Acq On : 25 Nov 2022 11:27
Operator : SY/MD
Sample : N5725-09DL 20X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BHB65DL

Quant Time: Nov 25 21:30:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 25 21:27:54 2022
Response via : Initial Calibration

