

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV112525\  
 Data File : VV039424.D  
 Acq On : 25 Nov 2025 12:17  
 Operator : SY/MD  
 Sample : VIBLK226  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VIBLK226

Quant Time: Nov 26 04:00:54 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVSIM112525.M  
 Quant Title : TRACE VOA SOM01.0  
 QLast Update : Wed Nov 26 03:57:32 2025  
 Response via : Initial Calibration

| Compound                      | R.T.           | QIon | Response            | Conc  | Units | Dev(Min) |
|-------------------------------|----------------|------|---------------------|-------|-------|----------|
| -----                         |                |      |                     |       |       |          |
| Internal Standards            |                |      |                     |       |       |          |
| 1) 1,4-Difluorobenzene        | 5.566          | 114  | 7221                | 0.500 | ug/L  | 0.00     |
| 5) Chlorobenzene-d5           | 8.785          | 117  | 6764                | 0.500 | ug/L  | -0.02    |
| 11) 1,4-Dichlorobenzene-d4    | 11.191         | 152  | 3503                | 0.500 | ug/L  | 0.00     |
| System Monitoring Compounds   |                |      |                     |       |       |          |
| 2) Vinyl Chloride-d3          | 1.295          | 65   | 3806                | 0.523 | ug/L  | 0.00     |
| Spiked Amount 0.500           | Range 40 - 130 |      | Recovery = 104.000% |       |       |          |
| 4) 1,2-Dichloroethane-d4      | 4.970          | 65   | 2030                | 0.447 | ug/L  | 0.00     |
| Spiked Amount 0.500           | Range 70 - 130 |      | Recovery = 90.000%  |       |       |          |
| 7) 1,2-Dichloropropane-d6     | 5.996          | 67   | 2243                | 0.523 | ug/L  | 0.00     |
| Spiked Amount 0.500           | Range 60 - 140 |      | Recovery = 104.000% |       |       |          |
| 8) Toluene-d8                 | 7.254          | 98   | 4400                | 0.519 | ug/L  | 0.00     |
| Spiked Amount 0.500           | Range 70 - 130 |      | Recovery = 104.000% |       |       |          |
| 10) 1,1,2,2-Tetrachloroeth... | 10.146         | 84   | 1553                | 0.516 | ug/L  | 0.00     |
| Spiked Amount 0.500           | Range 65 - 120 |      | Recovery = 104.000% |       |       |          |

Target Compounds Qvalue

-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\DATA\VV112525\  
Data File : VV039424.D  
Acq On : 25 Nov 2025 12:17  
Operator : SY/MD  
Sample : VIBLK226  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VIBLK226

Quant Time: Nov 26 04:00:54 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVSIM112525.M  
Quant Title : TRACE VOA SOM01.0  
QLast Update : Wed Nov 26 03:57:32 2025  
Response via : Initial Calibration

