

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112525\
 Data File : VV039425.D
 Acq On : 25 Nov 2025 12:38
 Operator : SY/MD
 Sample : VSTD0.0560
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.0260

Quant Time: Nov 26 03:50:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM112525.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Nov 26 03:49:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	6379	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	6359	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	3249	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.294	65	316	0.052	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.017	65	246	0.069	ug/L	0.05
7) 1,2-Dichloropropane-d6	6.019	67	212	0.059	ug/L	0.02
8) Toluene-d8	7.284	98	394	0.049	ug/L	0.03
10) 1,1,2,2-Tetrachloroeth...	10.146	84	147	0.057	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.294	62	434	0.040	ug/L #	24
6) Trichloroethene	5.876	95	273	0.049	ug/L	91
9) 1,2-Dibromoethane	8.321	107	202	0.060	ug/L #	83
12) 1,2,3-Trichloropropane	10.214	75	149	0.046	ug/L #	84
13) 1,2-Dibromo-3-chloropr...	12.374	75	35	0.047	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112525\
Data File : VV039425.D
Acq On : 25 Nov 2025 12:38
Operator : SY/MD
Sample : VSTD0.0560
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.0260

Quant Time: Nov 26 03:50:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM112525.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Nov 26 03:49:02 2025
Response via : Initial Calibration

