

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112525\
 Data File : VV039427.D
 Acq On : 25 Nov 2025 13:55
 Operator : SY/MD
 Sample : VV1125WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK232

Quant Time: Nov 26 04:03:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM112525.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Nov 26 03:57:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	6417	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	6399	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	3239	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.295	65	3496	0.540	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	108.000%
4) 1,2-Dichloroethane-d4	4.970	65	1865	0.463	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	92.000%
7) 1,2-Dichloropropane-d6	5.996	67	2149	0.530	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	106.000%
8) Toluene-d8	7.254	98	4032	0.502	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	1462	0.513	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	102.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112525\
Data File : VV039427.D
Acq On : 25 Nov 2025 13:55
Operator : SY/MD
Sample : VV1125WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK232

Quant Time: Nov 26 04:03:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM112525.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Nov 26 03:57:32 2025
Response via : Initial Calibration

