

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112525\
 Data File : VV039419.D
 Acq On : 25 Nov 2025 09:16
 Operator : SY/MD
 Sample : VSTD0.0555
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.05255

Quant Time: Nov 26 03:49:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM112525.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Nov 26 03:49:02 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	7910	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.785	117	7434	0.500	ug/L	-0.02
11) 1,4-Dichlorobenzene-d4	11.191	152	3861	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.294	65	491	0.065	ug/L	0.00
4) 1,2-Dichloroethane-d4	4.993	65	411	0.094	ug/L	0.02
7) 1,2-Dichloropropane-d6	6.019	67	198	0.047	ug/L	0.02
8) Toluene-d8	7.269	98	579	0.062	ug/L	0.02
10) 1,1,2,2-Tetrachloroeth...	10.146	84	152	0.051	ug/L	0.00
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.294	62	445	0.033	ug/L #	33
6) Trichloroethene	5.852	95	278	0.043	ug/L	90
9) 1,2-Dibromoethane	8.305	107	176	0.045	ug/L #	82
12) 1,2,3-Trichloropropane	10.214	75	161	0.042	ug/L	89
13) 1,2-Dibromo-3-chloropr...	12.356	75	27	0.031	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112525\
Data File : VV039419.D
Acq On : 25 Nov 2025 09:16
Operator : SY/MD
Sample : VSTD0.0555
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.05255

Quant Time: Nov 26 03:49:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM112525.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Nov 26 03:49:02 2025
Response via : Initial Calibration

