

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112525\
 Data File : VV039426.D
 Acq On : 25 Nov 2025 13:12
 Operator : SY/MD
 Sample : VSTDICV0.5
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV261

Quant Time: Nov 26 04:01:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM112525.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Nov 26 03:57:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	7194	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.785	117	6740	0.500	ug/L	-0.02
11) 1,4-Dichlorobenzene-d4	11.191	152	3513	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.294	65	3881	0.535	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 106.000%			
4) 1,2-Dichloroethane-d4	4.970	65	2109	0.467	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 94.000%			
7) 1,2-Dichloropropane-d6	5.996	67	2233	0.523	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 104.000%			
8) Toluene-d8	7.254	98	4528	0.536	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 108.000%			
10) 1,1,2,2-Tetrachloroeth...	10.146	84	1620	0.540	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery = 108.000%			
Target Compounds						
						Qvalue
3) Vinyl chloride	1.294	62	4969	0.526	ug/L	99
6) Trichloroethene	5.852	95	3167	0.544	ug/L	99
9) 1,2-Dibromoethane	8.290	107	2019	0.518	ug/L	97
12) 1,2,3-Trichloropropane	10.197	75	1671	0.526	ug/L	97
13) 1,2-Dibromo-3-chloropr...	12.357	75	342	0.491	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112525\
Data File : VV039426.D
Acq On : 25 Nov 2025 13:12
Operator : SY/MD
Sample : VSTDICV0.5
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VICV261

Quant Time: Nov 26 04:01:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM112525.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Nov 26 03:57:32 2025
Response via : Initial Calibration

