

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV112618\
 Data File : VV008715.D
 Acq On : 26 Nov 2018 15:20
 Operator : SY/MD
 Sample : J6036-04
 Misc : 5.0 mL/MSVOA V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0927

Quant Time: Nov 27 03:51:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM112018WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 27 03:11:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	268759	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	246478	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	101917	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	83368	43.97	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.94%
7) Chloroethane-d5	1.57	69	65270	63.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	126.04%
11) 1,1-Dichloroethene-d2	2.13	63	115849	37.02	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.04%
21) 2-Butanone-d5	3.93	46	148322	95.98	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.98%
24) Chloroform-d	4.41	84	185656	48.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.70%
26) 1,2-Dichloroethane-d4	5.08	65	123827	50.41	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.82%
32) Benzene-d6	5.10	84	386351	50.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.78%
36) 1,2-Dichloropropane-d6	6.12	67	122955	51.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.60%
41) Toluene-d8	7.36	98	347230	50.46	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.92%
43) trans-1,3-Dichloropropene-	7.67	79	52840	49.29	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.58%
47) 2-Hexanone-d5	8.13	63	100333	92.59	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.59%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	156400	47.84	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.68%
64) 1,2-Dichlorobenzene-d4	11.68	152	121027	54.40	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV112618\
Data File : VV008715.D
Acq On : 26 Nov 2018 15:20
Operator : SY/MD
Sample : J6036-04
Misc : 5.0 mL/MSVOA V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0927

Quant Time: Nov 27 03:51:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM112018WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 27 03:11:05 2018
Response via : Initial Calibration

