

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV112619\
 Data File : VV013809.D
 Acq On : 27 Nov 2019 00:27
 Operator : SY/MD
 Sample : K6004-07
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 B0AE7

Quant Time: Nov 27 05:03:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Nov 27 03:09:16 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	322382	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	319866	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	140762	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	75734	3.80	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.00%
7) Chloroethane-d5	1.58	69	77240	4.56	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.20%
11) 1,1-Dichloroethene-d2	2.13	63	108495	3.23	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.60%
20) 2-Butanone-d5	3.93	46	231618	55.84	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.68%
24) Chloroform-d	4.40	84	155558	3.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.40%
26) 1,2-Dichloroethane-d4	5.08	65	95816	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.00%
32) Benzene-d6	5.10	84	341590	4.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.80%
36) 1,2-Dichloropropane-d6	6.12	67	109020	4.43	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.60%
41) Toluene-d8	7.36	98	261387	3.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	67.60%#
43) trans-1,3-Dichloropropene-	7.66	79	35514	3.71	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.20%
46) 2-Hexanone-d5	8.13	63	170181	48.20	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.40%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	89290	5.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	116950	4.71	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.20%

Target Compounds

					Ovalue
13) Acetone	2.19	43	7201	2.483 ug/L #	41
25) Chloroform	4.43	83	31259	0.844 ug/L	99
34) Trichloroethene	5.96	95	4630	0.214 ug/L	89
47) Tetrachloroethene	8.02	164	1711	0.091 ug/L #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV112619\
Data File : VV013809.D
Acq On : 27 Nov 2019 00:27
Operator : SY/MD
Sample : K6004-07
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
B0AE7

Quant Time: Nov 27 05:03:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Nov 27 03:09:16 2019
Response via : Initial Calibration

