

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV112719\
 Data File : VV013820.D
 Acq On : 27 Nov 2019 13:38
 Operator : SY/MD
 Sample : K6004-08
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 B0AF0

Quant Time: Nov 29 02:33:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Nov 28 02:51:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	335072	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	325459	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	145140	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	78018	3.76	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.20%
7) Chloroethane-d5	1.58	69	78443	4.45	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.00%
11) 1,1-Dichloroethene-d2	2.13	63	114321	3.27	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.40%
20) 2-Butanone-d5	3.96	46	268617	62.31	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	124.62%
24) Chloroform-d	4.40	84	182885	4.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.40%
26) 1,2-Dichloroethane-d4	5.08	65	106061	5.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.60%
32) Benzene-d6	5.10	84	371144	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.40%
36) 1,2-Dichloropropane-d6	6.11	67	118345	4.73	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.60%
41) Toluene-d8	7.36	98	288865	3.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.40%
43) trans-1,3-Dichloropropene-	7.66	79	41035	4.21	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.20%
46) 2-Hexanone-d5	8.14	63	195608	54.45	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.90%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	96742	5.57	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	111.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	134419	5.25	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.00%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV112719\
Data File : VV013820.D
Acq On : 27 Nov 2019 13:38
Operator : SY/MD
Sample : K6004-08
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
B0AF0

Quant Time: Nov 29 02:33:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Nov 28 02:51:59 2019
Response via : Initial Calibration

