

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV112818\
 Data File : VV008752.D
 Acq On : 28 Nov 2018 21:52
 Operator : SY/MD
 Sample : J6096-06
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A4154

Quant Time: Nov 29 04:12:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112818WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Nov 29 03:24:09 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	156438	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	145198	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	61760	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	42785	4.72	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.40%
7) Chloroethane-d5	1.59	69	34113	5.53	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	110.60%
11) 1,1-Dichloroethene-d2	2.14	63	62803	3.62	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.40%
20) 2-Butanone-d5	3.96	46	120315	49.25	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.50%
24) Chloroform-d	4.41	84	99168	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.40%
26) 1,2-Dichloroethane-d4	5.09	65	51231	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.60%
32) Benzene-d6	5.10	84	207558	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
36) 1,2-Dichloropropane-d6	6.12	67	62915	4.73	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.60%
41) Toluene-d8	7.36	98	184143	4.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.60%
43) trans-1,3-Dichloropropene-	7.67	79	19841	4.25	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.00%
46) 2-Hexanone-d5	8.14	63	77090	43.43	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.86%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	39339	4.53	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	63093	5.17	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.40%

Target Compounds					Ovalue
13) Acetone	2.22	43	8318	6.903 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV112818\
Data File : VV008752.D
Acq On : 28 Nov 2018 21:52
Operator : SY/MD
Sample : J6096-06
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A4154

Quant Time: Nov 29 04:12:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112818WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Nov 29 03:24:09 2018
Response via : Initial Calibration

