

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112822\
 Data File : VV029371.D
 Acq On : 28 Nov 2022 10:41
 Operator : SY/MD
 Sample : VV1128WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Nov 28 22:14:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111022.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 11 02:00:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.615	114	24376	0.500	ug/L	-0.02
5) Chlorobenzene-d5	8.848	117	23038	0.500	ug/L	-0.02
11) 1,4-Dichlorobenzene-d4	11.244	152	13708	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	8909	0.384	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	76.000%
4) 1,2-Dichloroethane-d4	5.019	65	6773	0.478	ug/L	-0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	96.000%
7) 1,2-Dichloropropane-d6	6.045	67	7897	0.419	ug/L	-0.02
Spiked Amount	0.500	Range	60 - 140	Recovery	=	84.000%
8) Toluene-d8	7.301	98	29277	0.551	ug/L	-0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	110.000%
10) 1,1,2,2-Tetrachloroeth...	10.199	84	5539	0.455	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	90.000%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112822\
Data File : VV029371.D
Acq On : 28 Nov 2022 10:41
Operator : SY/MD
Sample : VV1128WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Nov 28 22:14:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111022.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 11 02:00:02 2022
Response via : Initial Calibration

