

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112822\
 Data File : VV029373.D
 Acq On : 28 Nov 2022 12:45
 Operator : SY/MD
 Sample : VSTDCCC0.5EC
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5291

Quant Time: Nov 28 22:14:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111022.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 11 02:00:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.615	114	20625	0.500	ug/L	-0.02
5) Chlorobenzene-d5	8.848	117	19368	0.500	ug/L	-0.02
11) 1,4-Dichlorobenzene-d4	11.244	152	11536	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	6997	0.357	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	72.000%
4) 1,2-Dichloroethane-d4	5.019	65	5279	0.440	ug/L	-0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	88.000%
7) 1,2-Dichloropropane-d6	6.068	67	5884	0.371	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	74.000%
8) Toluene-d8	7.301	98	22549	0.505	ug/L	-0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	102.000%
10) 1,1,2,2-Tetrachloroeth...	10.199	84	4349	0.425	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	84.000%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.303	62	6200	0.340	ug/L	92
6) Trichloroethene	5.901	95	6195	0.428	ug/L	98
9) 1,2-Dibromoethane	8.353	107	3761	0.448	ug/L	95
12) 1,2,3-Trichloropropane	10.250	75	3177	0.351	ug/L	93
13) 1,2-Dibromo-3-chloropr...	12.409	75	649	0.422	ug/L #	57

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112822\
Data File : VV029373.D
Acq On : 28 Nov 2022 12:45
Operator : SY/MD
Sample : VSTDCCC0.5EC
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5291

Quant Time: Nov 28 22:14:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111022.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 11 02:00:02 2022
Response via : Initial Calibration

