

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV112918\
 Data File : VV008773.D
 Acq On : 29 Nov 2018 19:12
 Operator : SY/MD
 Sample : J6076-03DL 4X
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YALQ2DL

Quant Time: Nov 30 06:03:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 30 04:55:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	146178	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	134918	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	58658	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	49647	4.91	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.20%
7) Chloroethane-d5	1.59	69	33858	4.71	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.20%
11) 1,1-Dichloroethene-d2	2.14	63	67410	3.88	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.60%
20) 2-Butanone-d5	3.96	46	112815	49.81	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.62%
24) Chloroform-d	4.41	84	97191	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.20%
26) 1,2-Dichloroethane-d4	5.09	65	48189	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.60%
32) Benzene-d6	5.10	84	207802	5.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.00%
36) 1,2-Dichloropropane-d6	6.12	67	59742	4.83	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.60%
41) Toluene-d8	7.36	98	179620	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.40%
43) trans-1,3-Dichloropropene-	7.67	79	18951	4.57	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.40%
46) 2-Hexanone-d5	8.14	63	79626	46.09	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.18%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	37326	4.77	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	61664	5.11	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.20%

Target Compounds

					Qvalue
2) Dichlorodifluoromethane	1.14	85	13719	1.061 ug/L	96
9) Trichlorofluoromethane	1.78	101	4220	0.293 ug/L	93
34) Trichloroethene	5.96	95	3802	0.338 ug/L	89
47) Tetrachloroethene	8.02	164	67440	7.752 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV112918\
Data File : VV008773.D
Acq On : 29 Nov 2018 19:12
Operator : SY/MD
Sample : J6076-03DL 4X
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YALQ2DL

Quant Time: Nov 30 06:03:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 30 04:55:05 2018
Response via : Initial Calibration

