

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV112918\
 Data File : VV008769.D
 Acq On : 29 Nov 2018 17:31
 Operator : SY/MD
 Sample : J6096-12
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Nov 30 05:46:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 30 04:55:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	169297	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	157663	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	70587	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	42618	3.64	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.80%
7) Chloroethane-d5	1.59	69	32145	3.86	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	77.20%
11) 1,1-Dichloroethene-d2	2.13	63	59227	2.94	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	58.80%#
20) 2-Butanone-d5	3.98	46	103736	39.54	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	79.08%
24) Chloroform-d	4.41	84	86236	3.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	74.40%
26) 1,2-Dichloroethane-d4	5.09	65	43580	3.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.60%
32) Benzene-d6	5.10	84	186571	3.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.60%
36) 1,2-Dichloropropane-d6	6.12	67	54468	3.76	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	75.20%
41) Toluene-d8	7.36	98	163055	3.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.40%
43) trans-1,3-Dichloropropene-	7.67	79	17747	3.66	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.20%
46) 2-Hexanone-d5	8.14	63	74065	36.69	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	73.38%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	32460	3.55	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	71.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	56009	3.86	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	77.20%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV112918\
Data File : VV008769.D
Acq On : 29 Nov 2018 17:31
Operator : SY/MD
Sample : J6096-12
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Nov 30 05:46:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 30 04:55:05 2018
Response via : Initial Calibration

