



Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112921\
 Data File : VV023738.D
 Acq On : 29 Nov 2021 10:22
 Operator : SY/MD
 Sample : BFB242
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFB242

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112321WMA.M
 Title : TRACE VOA SFAM1.0
 Last Update : Wed Dec 01 01:21:06 2021



AutoFind: Scans 2809, 2810, 2811; Background Corrected with Scan 2796

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.5	17397	PASS
75	95	30	80	51.7	51267	PASS
95	95	100	100	100.0	99160	PASS
96	95	5	9	6.4	6381	PASS
173	174	0.00	2	1.0	924	PASS
174	95	50	120	94.1	93309	PASS
175	174	5	9	7.3	6778	PASS
176	174	95	105	95.4	88995	PASS
177	176	5	10	6.2	5508	PASS

m/z	Abundance
39.90	162.0
43.90	138.0

Instrument :
MSVOA_V
ClientSampleId :
BFB242

m/z	Abundance
36.00	976.0
37.10	4104.0
38.10	3965.0
39.10	1692.0
39.90	171.0
44.00	747.0
45.00	825.0
45.90	104.0
47.00	1086.0
48.00	742.0
49.00	3717.0
50.10	17408.0
51.10	5572.0
52.00	216.0
53.00	224.0
54.90	304.0
56.00	1484.0
57.10	2557.0
58.00	147.0
60.00	856.0
61.00	4883.0
62.10	5250.0
63.00	4491.0
64.00	686.0
65.00	699.0
66.00	120.0
67.00	279.0
68.00	9893.0
69.00	9838.0
70.00	678.0
72.10	479.0
73.00	4415.0
74.00	17488.0
75.10	51080.0
76.00	4337.0
77.10	669.0
77.80	382.0
78.90	3830.0
79.90	1168.0
81.00	4440.0
82.00	1032.0
85.90	223.0
87.00	4397.0
88.00	3792.0
91.00	469.0
92.00	3262.0
93.00	4582.0
94.00	12453.0
95.00	98968.0
96.10	6466.0
97.10	208.0
103.80	596.0
104.80	174.0
105.90	630.0
106.90	109.0
111.10	159.0
112.80	164.0
114.90	242.0
115.90	439.0
116.90	791.0
118.00	524.0
118.90	888.0
124.10	112.0
127.90	474.0
128.90	271.0
129.90	284.0
130.70	159.0
134.90	201.0
136.90	242.0
139.80	128.0
140.90	1270.0
141.90	220.0
142.90	1474.0
144.20	144.0
144.90	434.0
145.90	186.0
147.90	332.0
148.90	106.0
149.90	113.0
154.90	273.0
156.90	247.0
158.80	156.0
160.80	140.0
170.80	202.0
172.00	1069.0
173.20	1364.0
174.00	90096.0
175.00	6617.0
176.00	85536.0
177.00	5560.0
178.00	143.0

Instrument :
MSVOA_V
ClientSampleId :
BFB242

m/z	Abundance
36.00	931.0
37.10	4329.0
38.10	4130.0
39.10	1750.0
39.80	176.0
40.90	103.0
44.00	686.0
45.00	784.0
45.90	149.0
47.00	1129.0
48.00	726.0
49.00	4132.0
50.10	18064.0
51.10	5608.0
52.00	254.0
53.00	235.0
54.90	289.0
56.00	1608.0
57.00	2815.0
58.00	151.0
58.80	101.0
60.00	921.0
61.00	5082.0
62.10	5258.0
63.00	4429.0
64.00	782.0
65.00	800.0
67.00	299.0
68.00	10282.0
69.00	10349.0
70.00	736.0
72.00	524.0
73.00	4712.0
74.00	17808.0
75.10	53200.0
76.00	4320.0
77.00	640.0
77.90	401.0
78.90	3848.0
79.90	1336.0
81.00	4454.0
82.00	1050.0
82.90	117.0
86.00	206.0
87.00	4680.0
88.00	3610.0
91.00	493.0
92.00	3501.0
93.00	4826.0
94.10	13133.0
95.00	102504.0
96.10	6588.0
97.10	196.0
103.90	658.0
104.80	180.0
105.90	646.0
106.90	128.0
111.00	152.0
112.80	179.0
114.80	261.0
115.90	426.0
116.90	926.0
117.90	529.0
118.90	927.0
124.00	104.0
127.90	561.0
128.90	255.0
129.90	326.0
130.70	171.0
134.90	242.0
137.00	305.0
139.80	157.0
140.90	1337.0
141.90	255.0
142.90	1571.0
143.80	148.0
144.90	465.0
145.90	186.0
146.90	112.0
147.90	328.0
148.90	111.0
150.00	104.0
154.90	348.0
156.90	271.0
158.90	163.0
160.80	151.0
170.80	250.0
172.00	1063.0
174.00	96352.0
175.00	7036.0
176.00	92224.0
177.00	5674.0

Instrument :
MSVOA_V
ClientSampleId :
BFB242

178.00

162.0

Instrument :
MSVOA_V
ClientSampleId :
BFB242

m/z	Abundance
36.00	846.0
37.10	4208.0
38.10	3863.0
39.10	1579.0
39.80	160.0
40.90	104.0
44.00	613.0
45.00	637.0
46.00	218.0
47.00	1072.0
48.00	579.0
49.00	4004.0
50.10	16720.0
51.10	5162.0
52.00	247.0
53.00	229.0
55.00	233.0
56.00	1527.0
57.00	2655.0
58.00	137.0
58.80	101.0
60.00	933.0
61.00	4807.0
62.00	4846.0
63.10	4079.0
64.00	733.0
65.00	862.0
67.00	288.0
68.00	9710.0
69.10	10002.0
70.00	746.0
72.00	515.0
73.00	4583.0
74.00	16408.0
75.10	49520.0
76.00	3965.0
77.00	575.0
77.90	376.0
78.90	3646.0
80.00	1345.0
80.90	4051.0
81.90	942.0
82.90	175.0
86.10	204.0
87.00	4378.0
88.00	3203.0
91.00	491.0
92.00	3182.0
93.00	4520.0
94.10	12450.0
95.00	96008.0
96.00	6089.0
97.10	177.0
103.90	629.0
104.80	173.0
105.90	664.0
106.80	143.0
110.90	152.0
112.80	161.0
114.80	240.0
115.90	381.0
116.90	929.0
117.90	504.0
118.90	847.0
127.90	518.0
128.80	184.0
129.90	357.0
130.70	155.0
134.90	228.0
137.00	323.0
139.80	130.0
140.90	1251.0
141.90	213.0
142.90	1630.0
143.80	129.0
144.90	417.0
145.80	207.0
146.80	128.0
147.90	261.0
154.90	395.0
156.90	231.0
159.00	131.0
160.80	125.0
170.90	267.0
172.00	1040.0
173.20	1408.0
174.00	93480.0
175.00	6680.0
176.00	89224.0
177.00	5289.0
178.00	147.0

Instrument :
MSVOA_V
ClientSampleId :
BFB242