

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW112922\
 Data File : VW029380.D
 Acq On : 29 Nov 2022 15:38
 Operator : SY/MD
 Sample : N5810-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGCE1

Quant Time: Nov 30 00:20:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 30 00:19:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	363325	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	325443	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	165658	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	85501	3.824	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.400%
7) Chloroethane-d5	1.564	69	67872	4.147	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.000%
11) 1,1-Dichloroethene-d2	2.105	63	103470	2.892	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.800%#
20) 2-Butanone-d5	3.912	46	175989	47.051	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.100%
24) Chloroform-d	4.339	84	163636	4.100	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.000%
26) 1,2-Dichloroethane-d4	5.027	65	76251	4.342	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.800%
32) Benzene-d6	5.043	84	360641	4.010	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.200%
36) 1,2-Dichloropropane-d6	6.063	67	106707	4.222	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.400%
41) Toluene-d8	7.307	98	325884	3.839	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.800%
43) trans-1,3-Dichloroprop...	7.616	79	39890	4.094	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.800%
46) 2-Hexanone-d5	8.085	63	168738	48.580	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.160%
56) 1,1,2,2-Tetrachloroeth...	10.207	84	71470	4.583	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.600%
66) 1,2-Dichlorobenzene-d4	11.616	152	121070	4.184	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	83.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112922\
Data File : VV029380.D
Acq On : 29 Nov 2022 15:38
Operator : SY/MD
Sample : N5810-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGCE1

Quant Time: Nov 30 00:20:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 30 00:19:33 2022
Response via : Initial Calibration

