

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV113018\
 Data File : VV008805.D
 Acq On : 30 Nov 2018 11:49
 Operator : SY/MD
 Sample : VV1130WBL01
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK77

Quant Time: Dec 01 04:38:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Dec 01 03:50:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	173673	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	162552	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	73034	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	43406	3.62	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.40%
7) Chloroethane-d5	1.59	69	35101	4.11	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.20%
11) 1,1-Dichloroethene-d2	2.13	63	63425	3.07	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.40%
20) 2-Butanone-d5	3.98	46	103679	38.53	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	77.06%
24) Chloroform-d	4.41	84	97456	4.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.00%
26) 1,2-Dichloroethane-d4	5.09	65	48676	4.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.40%
32) Benzene-d6	5.10	84	206087	4.15	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.00%
36) 1,2-Dichloropropane-d6	6.12	67	60815	4.08	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.60%
41) Toluene-d8	7.36	98	182488	3.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.60%
43) trans-1,3-Dichloropropene-	7.67	79	18867	3.77	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.40%
46) 2-Hexanone-d5	8.14	63	85771	41.21	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.42%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	35434	3.76	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	75.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	65490	4.36	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	87.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV113018\
Data File : VV008805.D
Acq On : 30 Nov 2018 11:49
Operator : SY/MD
Sample : VV1130WBL01
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK77

Quant Time: Dec 01 04:38:08 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Dec 01 03:50:55 2018
Response via : Initial Calibration

