

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV113018\
 Data File : VV008808.D
 Acq On : 30 Nov 2018 13:12
 Operator : SY/MD
 Sample : J6076-18
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YALS9

Quant Time: Dec 01 04:58:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Dec 01 03:50:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	135741	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	126702	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	55889	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	47021	5.01	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.20%
7) Chloroethane-d5	1.59	69	33173	4.97	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.40%
11) 1,1-Dichloroethene-d2	2.14	63	67103	4.16	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.20%
20) 2-Butanone-d5	3.96	46	116573	55.42	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.84%
24) Chloroform-d	4.41	84	102287	5.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.20%
26) 1,2-Dichloroethane-d4	5.09	65	51788	5.82	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	116.40%
32) Benzene-d6	5.10	84	212406	5.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.80%
36) 1,2-Dichloropropane-d6	6.12	67	62833	5.40	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.00%
41) Toluene-d8	7.36	98	186856	5.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.60%
43) trans-1,3-Dichloropropene-	7.67	79	20211	5.19	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	103.80%
46) 2-Hexanone-d5	8.14	63	81617	50.31	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.62%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	36943	5.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	65891	5.73	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.60%

Target Compounds					Ovalue
13) Acetone	2.21	43	16659	15.892 ug/L	96
47) Tetrachloroethene	8.02	164	4873	0.596 ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV113018\
Data File : VV008808.D
Acq On : 30 Nov 2018 13:12
Operator : SY/MD
Sample : J6076-18
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YALS9

Quant Time: Dec 01 04:58:09 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Dec 01 03:50:55 2018
Response via : Initial Calibration

