


Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV113018\
Data File : VV008803.D
Acq On : 30 Nov 2018 09:37
Operator : SY/MD
Sample : BFB53
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFB53

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR112918WMA.M
Title : TRACE VOA SOM01.0
Last Update : Sat Dec 01 03:50:55 2018



AutoFind: Scans 2808, 2809, 2810; Background Corrected with Scan 2795

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.3	14712	PASS
75	95	30	80	48.3	41093	PASS
95	95	100	100	100.0	85050	PASS
96	95	5	9	6.4	5401	PASS
173	174	0.00	2	1.5	995	PASS
174	95	50	120	78.1	66448	PASS
175	174	5	9	8.1	5373	PASS
176	174	95	101	98.6	65528	PASS
177	176	5	9	6.5	4264	PASS

m/z	Abundance
35.90	684.0
37.00	3465.0
38.00	3083.0
39.00	1320.0
39.90	199.0
44.10	629.0
45.00	525.0
46.90	778.0
47.90	567.0
49.00	2799.0
50.00	14301.0
51.00	4639.0
52.10	177.0
54.80	168.0
55.20	148.0
55.90	1067.0
57.00	2132.0
60.00	679.0
61.00	3363.0
62.00	3588.0
63.00	2827.0
64.00	404.0
65.00	335.0
66.90	196.0
68.00	7039.0
69.00	7696.0
70.00	583.0
72.00	516.0
73.00	3380.0
74.00	13423.0
75.00	39552.0
76.00	3443.0
77.00	512.0
77.90	315.0
78.90	2498.0
79.90	792.0
80.90	2407.0
81.90	521.0
85.90	175.0
87.00	3161.0
87.90	3022.0
90.90	260.0
91.90	2511.0
93.00	3459.0
94.00	9599.0
95.00	80760.0
96.00	5240.0
97.10	224.0
103.80	357.0
106.00	281.0
112.90	105.0
115.90	298.0
116.80	502.0
117.90	406.0
118.90	500.0
127.90	340.0
128.90	168.0
130.00	257.0
134.90	147.0
136.90	115.0
140.90	876.0
142.00	117.0
142.80	962.0
144.90	208.0
145.90	181.0
147.80	161.0
149.80	128.0
154.90	199.0
157.00	104.0
158.80	108.0
160.90	144.0
171.00	126.0
171.90	705.0
173.00	891.0
174.00	61904.0
175.00	5118.0
176.00	61296.0
177.00	4061.0

Instrument :
MSVOA_V
ClientSampleId :
BFB53

Scan 2795 (10.077 min): VV008803.D
BFB53

m/z	Abundance
40.00	243.0
43.90	335.0

Instrument :
MSVOA_V
ClientSampleId :
BFB53

m/z	Abundance
35.90	657.0
37.00	3223.0
38.00	3138.0
39.00	1306.0
39.90	309.0
44.10	539.0
45.00	551.0
47.00	898.0
47.90	494.0
49.00	3265.0
50.00	14618.0
51.00	4784.0
51.90	184.0
54.90	182.0
56.00	1244.0
57.00	2104.0
59.90	766.0
61.00	3640.0
62.00	3703.0
63.00	2869.0
64.00	550.0
65.00	432.0
66.90	223.0
68.00	7935.0
69.00	7985.0
69.90	592.0
72.00	567.0
73.00	3438.0
74.00	13801.0
75.00	41256.0
76.00	3442.0
76.90	542.0
77.90	319.0
78.90	2378.0
79.90	821.0
80.90	2383.0
81.90	613.0
85.90	202.0
86.90	3464.0
88.00	3635.0
90.90	335.0
92.00	2388.0
93.00	3466.0
94.00	10557.0
95.00	86320.0
96.00	5383.0
97.10	170.0
103.90	334.0
105.00	129.0
105.80	237.0
115.80	302.0
116.80	486.0
117.90	369.0
118.90	564.0
127.90	328.0
128.70	126.0
130.00	383.0
134.80	198.0
136.80	136.0
140.90	1051.0
142.80	1096.0
144.90	214.0
145.90	173.0
147.70	130.0
148.00	123.0
149.90	132.0
152.70	112.0
155.00	256.0
157.10	119.0
161.00	128.0
172.00	814.0
172.90	1078.0
174.00	68744.0
175.00	5465.0
176.00	67560.0
177.00	4316.0

Instrument :
MSVOA_V
ClientSampleId :
BFB53

m/z	Abundance
35.90	726.0
37.00	3501.0
38.00	3310.0
39.00	1352.0
39.90	260.0
44.10	619.0
45.00	553.0
46.90	908.0
47.90	580.0
49.00	3197.0
50.00	15217.0
51.00	4985.0
52.00	163.0
54.80	184.0
55.90	1192.0
57.00	2244.0
60.00	780.0
61.00	3674.0
62.00	3901.0
63.00	2907.0
64.00	517.0
64.90	402.0
66.90	226.0
68.00	7781.0
69.00	8270.0
70.00	617.0
72.00	580.0
73.00	3606.0
74.00	14462.0
75.00	42472.0
76.00	3572.0
76.90	523.0
77.90	349.0
78.90	2526.0
79.90	836.0
80.90	2545.0
81.90	556.0
85.90	213.0
86.90	3582.0
87.90	3456.0
90.90	338.0
92.00	2511.0
93.00	3590.0
94.00	10619.0
95.00	88072.0
96.00	5582.0
97.10	226.0
103.80	375.0
105.80	262.0
115.80	294.0
116.80	539.0
117.90	426.0
118.90	561.0
127.90	365.0
128.80	164.0
130.00	335.0
134.80	179.0
136.90	146.0
140.90	1056.0
142.00	122.0
142.80	1114.0
144.90	241.0
145.90	197.0
147.70	145.0
148.00	123.0
149.90	131.0
154.90	256.0
157.10	104.0
161.00	146.0
172.00	798.0
173.00	1017.0
174.00	68696.0
175.00	5538.0
176.00	67728.0
177.00	4417.0

Instrument :
MSVOA_V
ClientSampleId :
BFB53