




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV113021\
Data File : VV023746.D
Acq On : 30 Nov 2021 10:27
Operator : SY/MD
Sample : VV1130WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Dec 01 01:21:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 01 01:21:06 2021
Response via : Initial Calibration



TIC: VV023746.D\data.ms

(20) 2-Butanone-d5 (S)




3.896min (-0.016) 32.85 ug/L

response 44957

Ion	Exp%	Act%
46.10	100.00	100.00
77.10	9.40	19.48#
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV113021\
Data File : VV023746.D
Acq On : 30 Nov 2021 10:27
Operator : SY/MD
Sample : VV1130WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Dec 01 01:21:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 01 01:21:06 2021
Response via : Initial Calibration



TIC: VV023746.D\data.ms

(20) 2-Butanone-d5 (S)**3.896min (-0.016) 39.74 ug/L m****response 54383**

Ion	Exp%	Act%
46.10	100.00	100.00
77.10	9.40	16.10#
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV113021\
 Data File : VV023746.D
 Acq On : 30 Nov 2021 10:27
 Operator : SY/MD
 Sample : VV1130WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Quant Time: Dec 01 01:21:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 01 01:21:06 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	138673	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	134903	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	63974	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	44467	3.906	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	78.200%	
7) Chloroethane-d5	1.564	69	36601	4.090	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	81.800%	
11) 1,1-Dichloroethene-d2	2.105	63	64933	3.237	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	64.800%	
20) 2-Butanone-d5	3.896	46	54383m	39.738	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	79.480%	
24) Chloroform-d	4.342	84	82440	4.159	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	83.200%	
26) 1,2-Dichloroethane-d4	5.027	65	41646	4.497	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.000%	
32) Benzene-d6	5.043	84	169810	4.621	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.400%	
36) 1,2-Dichloropropane-d6	6.066	67	47752	4.635	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	92.800%	
41) Toluene-d8	7.313	98	147985	4.310	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	86.200%	
43) trans-1,3-Dichloroprop...	7.619	79	17310	4.168	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	83.400%	
46) 2-Hexanone-d5	8.088	63	54372	39.408	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	78.820%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	30040	4.052	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	81.000%	
66) 1,2-Dichlorobenzene-d4	11.625	152	56867	5.028	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	100.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV113021\
Data File : VV023746.D
Acq On : 30 Nov 2021 10:27
Operator : SY/MD
Sample : VV1130WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Dec 01 01:21:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 01 01:21:06 2021
Response via : Initial Calibration

