

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW113021\
 Data File : VW023746.D
 Acq On : 30 Nov 2021 10:27
 Operator : SY/MD
 Sample : VW1130WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK265

Quant Time: Dec 01 01:21:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 01 01:21:06 2021
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :John
 Carlone
 12/01/2021
 Supervised By :Mahesh
 Dadoda
 12/01/2021

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	138673	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	134903	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	63974	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	44467	3.906	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.200%
7) Chloroethane-d5	1.564	69	36601	4.090	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.800%
11) 1,1-Dichloroethene-d2	2.105	63	64933	3.237	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.800%
20) 2-Butanone-d5	3.896	46	54383m	39.738	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	79.480%
24) Chloroform-d	4.342	84	82440	4.159	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.200%
26) 1,2-Dichloroethane-d4	5.027	65	41646	4.497	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.000%
32) Benzene-d6	5.043	84	169810	4.621	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.400%
36) 1,2-Dichloropropane-d6	6.066	67	47752	4.635	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.800%
41) Toluene-d8	7.313	98	147985	4.310	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.200%
43) trans-1,3-Dichloroprop...	7.619	79	17310	4.168	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.400%
46) 2-Hexanone-d5	8.088	63	54372	39.408	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	78.820%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	30040	4.052	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	81.000%
66) 1,2-Dichlorobenzene-d4	11.625	152	56867	5.028	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV113021\
Data File : VV023746.D
Acq On : 30 Nov 2021 10:27
Operator : SY/MD
Sample : VV1130WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK265

Quant Time: Dec 01 01:21:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 01 01:21:06 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John

Carlone

12/01/2021

Supervised By :Mahesh

Dadoda

12/01/2021

