

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120123\
 Data File : VV033081.D
 Acq On : 01 Dec 2023 16:06
 Operator : SY/MD
 Sample : VV1201MBL01
 Misc : 5.0G/5.0mL/100uL/5mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK229

Quant Time: Dec 04 04:22:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Dec 04 04:21:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	357574	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	353675	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	192216	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	109207	51.089	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 102.180%	
7) Chloroethane-d5	1.532	69	74504	50.023	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 100.040%	
11) 1,1-Dichloroethene-d2	2.059	65	41363	50.865	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 101.740%	
21) 2-Butanone-d5	3.780	46	102891	96.912	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 96.910%	
24) Chloroform-d	4.249	84	217542	47.955	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 95.920%	
26) 1,2-Dichloroethane-d4	4.940	65	139003	49.285	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 98.580%	
32) Benzene-d6	4.960	84	421983	51.016	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 102.040%	
36) 1,2-Dichloropropane-d6	5.989	67	115512	51.938	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 103.880%	
41) Toluene-d8	7.243	98	399252	50.598	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 101.200%	
43) trans-1,3-Dichloroprop...	7.551	79	67392	47.919	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 95.840%	
47) 2-Hexanone-d5	8.024	63	82933	92.820	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 92.820%	
56) 1,1,2,2-Tetrachloroeth...	10.152	84	143506	45.391	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 90.780%	
66) 1,2-Dichlorobenzene-d4	11.561	152	177214	51.817	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 103.640%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120123\
Data File : VV033081.D
Acq On : 01 Dec 2023 16:06
Operator : SY/MD
Sample : VV1201MBL01
Misc : 5.0G/5.0mL/100uL/5mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK229

Quant Time: Dec 04 04:22:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Dec 04 04:21:32 2023
Response via : Initial Calibration

