

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW120123\
 Data File : VW033085.D
 Acq On : 01 Dec 2023 17:39
 Operator : SY/MD
 Sample : VIBLK259
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK259

Quant Time: Dec 04 04:23:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Dec 04 04:21:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	316349	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	316331	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	169837	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	93236	49.302	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	98.600%
7) Chloroethane-d5	1.532	69	68427	51.930	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	103.860%
11) 1,1-Dichloroethene-d2	2.060	65	36932	51.335	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	102.660%
21) 2-Butanone-d5	3.780	46	93386	99.422	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.420%
24) Chloroform-d	4.249	84	208514	51.955	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.920%
26) 1,2-Dichloroethane-d4	4.944	65	141313	56.634	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	113.260%
32) Benzene-d6	4.960	84	406355	54.926	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	109.860%
36) 1,2-Dichloropropane-d6	5.989	67	112704	56.657	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	113.320%
41) Toluene-d8	7.243	98	377440	53.480	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.960%
43) trans-1,3-Dichloroprop...	7.551	79	65764	52.282	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	104.560%
47) 2-Hexanone-d5	8.024	63	78769	98.568	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.570%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	146511	51.812	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	103.620%
66) 1,2-Dichlorobenzene-d4	11.561	152	169031	55.937	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.880%
Target Compounds						
51) Chlorobenzene	8.818	112	5593	0.880	ug/L	89
65) 1,4-Dichlorobenzene	11.210	146	3467	0.638	ug/L #	81
70) 1,2,4-trichlorobenzene	13.204	180	5262	1.359	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120123\
Data File : VV033085.D
Acq On : 01 Dec 2023 17:39
Operator : SY/MD
Sample : VIBLK259
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK259

Quant Time: Dec 04 04:23:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM120123WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Dec 04 04:21:32 2023
Response via : Initial Calibration

