

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV120220\
 Data File : VV019444.D
 Acq On : 02 Dec 2020 13:01
 Operator : SY/MD
 Sample : VV1202WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK90

Quant Time: Dec 03 05:55:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM112320WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 03 05:50:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	368697	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	374359	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	179930	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	137504	37.34	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	74.68%
7) Chloroethane-d5	1.58	69	125500	41.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.74%
11) 1,1-Dichloroethene-d2	2.12	63	217505	33.18	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.36%
21) 2-Butanone-d5	3.89	46	228308	101.73	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.73%
24) Chloroform-d	4.37	84	264089	44.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.72%
26) 1,2-Dichloroethane-d4	5.05	65	183054	46.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.96%
32) Benzene-d6	5.07	84	504680	45.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.20%
36) 1,2-Dichloropropane-d6	6.09	67	175425	47.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.28%
41) Toluene-d8	7.33	98	423145	42.67	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.34%
43) trans-1,3-Dichloropropene-	7.64	79	81842	46.01	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.02%
47) 2-Hexanone-d5	8.10	63	122563	84.47	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.47%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	206543	43.33	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.66%
64) 1,2-Dichlorobenzene-d4	11.65	152	174665	49.99	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.98%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV120220\
Data File : VV019444.D
Acq On : 02 Dec 2020 13:01
Operator : SY/MD
Sample : VV1202WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK90

Quant Time: Dec 03 05:55:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM112320WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 03 05:50:19 2020
Response via : Initial Calibration

