

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW120222\
 Data File : VW029388.D
 Acq On : 02 Dec 2022 12:37
 Operator : SY/MD
 Sample : VW1202WBL02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK281

Quant Time: Dec 02 21:18:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 02 21:17:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	258280	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	233446	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	116465	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	64478	4.057	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.200%
7) Chloroethane-d5	1.565	69	54445	4.679	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.600%
11) 1,1-Dichloroethene-d2	2.105	63	81542	3.206	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.200%
20) 2-Butanone-d5	3.886	46	148020	55.668	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.340%
24) Chloroform-d	4.339	84	143749	5.067	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.400%
26) 1,2-Dichloroethane-d4	5.024	65	64735	5.186	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.800%
32) Benzene-d6	5.043	84	300147	4.652	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.000%
36) 1,2-Dichloropropane-d6	6.063	67	92155	5.083	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.600%
41) Toluene-d8	7.307	98	272004	4.467	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.400%
43) trans-1,3-Dichloroprop...	7.616	79	33905	4.851	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.000%
46) 2-Hexanone-d5	8.082	63	138375	55.538	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	111.080%
56) 1,1,2,2-Tetrachloroeth...	10.207	84	64991	5.810	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	116.200%
66) 1,2-Dichlorobenzene-d4	11.616	152	101334	4.981	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV120222\
Data File : VV029388.D
Acq On : 02 Dec 2022 12:37
Operator : SY/MD
Sample : VV1202WBL02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK281

Quant Time: Dec 02 21:18:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 02 21:17:55 2022
Response via : Initial Calibration

