

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV120319\
 Data File : VV013895.D
 Acq On : 03 Dec 2019 15:55
 Operator : SY/MD
 Sample : K6028-14
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0AB0

Quant Time: Dec 04 03:41:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120219WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 04 01:11:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	564166	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	524407	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	216559	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	179397	54.98	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	109.96%
7) Chloroethane-d5	1.57	69	199845	58.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	117.36%
11) 1,1-Dichloroethene-d2	2.12	63	304863	42.25	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.50%
21) 2-Butanone-d5	3.92	46	273714	119.18	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	119.18%
24) Chloroform-d	4.39	84	416623	60.80	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	121.60%
26) 1,2-Dichloroethane-d4	5.07	65	269150	61.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	123.98%
32) Benzene-d6	5.09	84	821710	60.47	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	120.94%
36) 1,2-Dichloropropane-d6	6.11	67	257935	63.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	126.48%#
41) Toluene-d8	7.35	98	724092	58.45	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	116.90%
43) trans-1,3-Dichloropropene-	7.66	79	114096	57.48	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	114.96%
47) 2-Hexanone-d5	8.13	63	143322	94.10	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.10%
57) 1,1,2,2-Tetrachloroethane-	10.25	84	338374	57.29	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	114.58%
64) 1,2-Dichlorobenzene-d4	11.67	152	294391	69.21	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	138.42%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV120319\
Data File : VV013895.D
Acq On : 03 Dec 2019 15:55
Operator : SY/MD
Sample : K6028-14
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0AB0

Quant Time: Dec 04 03:41:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM120219WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 04 01:11:22 2019
Response via : Initial Calibration

