

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120324\
 Data File : VV038398.D
 Acq On : 03 Dec 2024 09:35
 Operator : SY/MD
 Sample : VSTD0.127
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.1227

Quant Time: Dec 03 20:54:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120324.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Dec 03 20:54:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	9258	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	9379	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	4501	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	840	0.091	ug/L	0.00
4) 1,2-Dichloroethane-d4	4.993	65	633	0.108	ug/L	0.02
7) 1,2-Dichloropropane-d6	6.019	67	820	0.106	ug/L	0.02
8) Toluene-d8	7.269	98	1958	0.101	ug/L	0.02
10) 1,1,2,2-Tetrachloroeth...	10.146	84	652	0.106	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.280	62	1078	0.111	ug/L	91
6) Trichloroethene	5.852	95	797	0.114	ug/L	96
9) 1,2-Dibromoethane	8.321	107	505	0.106	ug/L #	90
12) 1,2,3-Trichloropropane	10.214	75	459	0.115	ug/L	97
13) 1,2-Dibromo-3-chloropr...	12.374	75	76	0.099	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120324\
Data File : VV038398.D
Acq On : 03 Dec 2024 09:35
Operator : SY/MD
Sample : VSTD0.127
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.1227

Quant Time: Dec 03 20:54:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120324.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Dec 03 20:54:04 2024
Response via : Initial Calibration

