

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120324\
 Data File : VV038400.D
 Acq On : 03 Dec 2024 10:30
 Operator : SY/MD
 Sample : VSTD00129
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD001229

Quant Time: Dec 03 20:54:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120324.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Dec 03 20:54:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	9383	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	9493	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	4405	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	8359	0.889	ug/L	0.00
4) 1,2-Dichloroethane-d4	4.946	65	6722	1.128	ug/L	-0.02
7) 1,2-Dichloropropane-d6	5.995	67	8985	1.144	ug/L	0.00
8) Toluene-d8	7.238	98	20224	1.036	ug/L	-0.02
10) 1,1,2,2-Tetrachloroeth...	10.146	84	7012	1.129	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.280	62	10601	1.076	ug/L	98
6) Trichloroethene	5.828	95	7652	1.081	ug/L	92
9) 1,2-Dibromoethane	8.290	107	5153	1.073	ug/L #	97
12) 1,2,3-Trichloropropane	10.197	75	4533	1.159	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.357	75	804	1.069	ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120324\
Data File : VV038400.D
Acq On : 03 Dec 2024 10:30
Operator : SY/MD
Sample : VSTD00129
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD001229

Quant Time: Dec 03 20:54:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120324.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Dec 03 20:54:04 2024
Response via : Initial Calibration

