

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120324\
 Data File : VV038402.D
 Acq On : 03 Dec 2024 11:32
 Operator : SY/MD
 Sample : VSTDICV0.5
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV231

Quant Time: Dec 03 21:47:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120324.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Dec 03 21:43:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	9105	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	9163	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	4442	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.273	65	4170	0.496	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 100.000%	
4) 1,2-Dichloroethane-d4	4.970	65	3356	0.520	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 104.000%	
7) 1,2-Dichloropropane-d6	5.996	67	4255	0.496	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	= 100.000%	
8) Toluene-d8	7.254	98	9651	0.480	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 96.000%	
10) 1,1,2,2-Tetrachloroeth...	10.146	84	3339	0.501	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 100.000%	
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.280	62	5445	0.505	ug/L	97
6) Trichloroethene	5.853	95	3844	0.494	ug/L	99
9) 1,2-Dibromoethane	8.305	107	2534	0.505	ug/L	99
12) 1,2,3-Trichloropropane	10.197	75	2253	0.487	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.357	75	392	0.494	ug/L #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120324\
Data File : VV038402.D
Acq On : 03 Dec 2024 11:32
Operator : SY/MD
Sample : VSTDICV0.5
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VICV231

Quant Time: Dec 03 21:47:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120324.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Dec 03 21:43:44 2024
Response via : Initial Calibration

