

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120324\
 Data File : VV038403.D
 Acq On : 03 Dec 2024 12:58
 Operator : SY/MD
 Sample : VV1203WBL01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK214

Quant Time: Dec 03 21:47:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120324.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Dec 03 21:43:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	8501	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	8556	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	4078	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.273	65	4194	0.534	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 106.000%	
4) 1,2-Dichloroethane-d4	4.969	65	3426	0.568	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 114.000%	
7) 1,2-Dichloropropane-d6	5.995	67	4308	0.537	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	= 108.000%	
8) Toluene-d8	7.254	98	9350	0.498	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 100.000%	
10) 1,1,2,2-Tetrachloroeth...	10.146	84	3435	0.552	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 110.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120324\
Data File : VV038403.D
Acq On : 03 Dec 2024 12:58
Operator : SY/MD
Sample : VV1203WBL01
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK214

Quant Time: Dec 03 21:47:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120324.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Dec 03 21:43:44 2024
Response via : Initial Calibration

